

SPECTRUM™ AFO SYSTEM

MID-CALF



Spectrum 100

Free Motion
Split Upright

When mild control is needed for medial/lateral instability. This functionally efficient completely custom AFO offers split uprights, mechanical ankle axis, EVA topcover, balanced footplate, low-profile trimlines. Removable interface pads are designed to be washable and add needed comfort. Requires mid-calf cast.



Spectrum 200

Free Motion
Solid Upright

When moderate control is needed for medial/lateral instability. Custom AFO with balanced footplates, mechanical ankle axis and custom uprights that connect to posterior bar adding more stability and control. Removable interface pads are designed to be washable and add needed comfort. Requires mid-calf cast.



Spectrum 300

Dorsi-Assist
Solid Upright

When moderate control is needed for medial/lateral instability. With the added feature of Tamarack dorsi-assist ankle joints. Designed for patients with drop foot during the swing phase of gait. Removable interface pads are designed to be washable and add needed comfort. Requires mid-calf cast.

ORDER INFORMATION

Customer Information

Date	Company	PO#	Email
Contact Person		Phone	Fax
Address	City	State	Zip
Ship Via/Shipping Instructions		Date Required	

Protected Health Information

Patient Name/ID	Sex	Height	Weight	Age
Diagnosis/Notes				

Custom Fabrication Hotline 877-737-8444
Fax 877-737-8445 | orthomerica.com/spectrum
6333 N Orange Blossom Trl, Orlando FL 32810

BILL TO:

Cascade Orthopedic Supply
2638 Aztec Drive • Chico, CA 95928

ORTHOMERICA®

©2015 Orthomerica Products, Inc. All Rights Reserved. 505-0000-160 Rev. E

IMPORTANT!

All information must be completed or your order will be put on hold.

Unspecified options will default to those in **bold**.

Items marked with * will incur additional charges.

MID-CALF DESIGN

- ☐ **Spectrum 100** ☐ **Spectrum 200** ☐ **Spectrum 300**
- Free Motion Free Motion Dorsi-Assist
Split Upright Solid Upright Solid Upright

Color

- ☐ **Black** ☐ White

Side

- ☐ Left ☐ Right ☐ Bilateral

Crepe Heel Post

- ☐ **None** ☐ Heel Only ☐ Heel to Metatarsal*

Finished Heel Post Position

- ☐ **Neutral** ☐ Medial _____ ☐ Lateral _____

Crepe Forefoot Post Position

- ☐ **None** ☐ Medial* _____ ☐ Lateral* _____ ☐ Metatarsal to end of toes* _____

Foot Plate Liner Options

- ☐ **Standard (EVA)** ☐ Poron ☐ Plastazote® ☐ Bi-laminate Poron Plastazote* ☐ Add Navicular Pad*

Foot Plate Plastic Length

- ☐ **To Mets** ☐ To Sulcus ☐ To Toes

Foot Plate Liner Length

- ☐ **To Sulcus** ☐ To Mets ☐ To Toes

Heel Cup Height

- ☐ **35mm (1 3/8") Standard** ☐ Other _____

Heel Lift Options

- ☐ Heel lift _____ inches *

Plastic Trim Options

- ☐ High Medial Wall to 1st Metatarsal ☐ High Lateral Wall to 5th Metatarsal ☐ Low Side Walls

MODIFICATIONS

Modification Options

- ☐ Navicular Relief ☐ Styloid 5th Met Relief ☐ Metatarsal Rise ☐ Other _____

Arch Options

- ☐ **Standard (to Cast)** ☐ Increase Arch Height by _____ ☐ Decrease Arch Height by _____

Cast Correction

Ankle Alignment


- ☐ **90°** ☐ _____ ° ☐ Dorsiflexion ☐ Plantarflexion ☐ Do Not Correct

Hindfoot Subtalar Alignment

- ☐ **Neutral** ☐ Do Not Correct

Forefoot Alignment

Circle drawing below to indicate finished forefoot alignment.




Include Met Height to Floor in Inches

Measurements



Standard brace height is 9"
Depending on request additional charges may apply



Special Instructions

Custom Fabrication Hotline 877-737-8444

Fax 877-737-8445 | orthomerica.com/spectrum

6333 N Orange Blossom Trl, Orlando FL 32810

ORTHOMERICA®

©2015 Orthomerica Products, Inc. All Rights Reserved. 505-0000-160 Rev. E